

# Diabetes Technology—Continuous Subcutaneous Insulin Infusion Therapy and Continuous Glucose Monitoring in Adults: An Endocrine Society Clinical Practice Guideline

Authors: Anne L. Peters (chair), Andrew J. Ahmann, Tadej Battelino, Alison Evert, Irl B. Hirsch, M. Hassan Murad, William E. Winter, Howard Wolpert

Keck School of Medicine, University of Southern California (A.L.P.), Los Angeles, CA 90033; Harold Schnitzer Diabetes Health Center, Oregon Health and Science University (A.J.A.), Portland, OR 97239; Department of Medicine, University of Ljubljana and University Children's Hospital (T.B.), 1104 Ljubljana, Slovenia; Endocrine and Diabetes Care Center, University of Washington Medical Center (A.E.), Seattle, WA 98195; Endocrine and Diabetes Care Center, University of Washington Medical Center (I.B.H.), Seattle, WA 98195; Mayo Clinic Evidence-based Practice Center, (M.H.M.), Rochester, MN 55905; UF Diabetes Institute, University of Florida (W.E.W.), Gainesville, FL 32611; Joslin Diabetes Center, Harvard Medical School (H.W.), Boston, MA 02115

**Objective:** To formulate clinical practice guidelines for the use of continuous glucose monitoring and continuous subcutaneous insulin infusion in adults with diabetes.

**Participants:** The participants include an Endocrine Society-appointed Task Force of seven experts, a methodologist, and a medical writer. The American Association for Clinical Chemistry, the American Association of Diabetes Educators, and the European Society of Endocrinology co-sponsored this guideline.

**Evidence:** The Task Force developed this evidence-based guideline using the Grading of Recommendations, Assessment, Development, and Evaluation system to describe the strength of recommendations and the quality of evidence. The Task Force commissioned one systematic review and used the best available evidence from other published systematic reviews and individual studies.

**Consensus Process:** One group meeting, several conference calls, and e-mail communications enabled consensus. Committees and members of the Endocrine Society, the American Association for Clinical Chemistry, the American Association of Diabetes Educators, and the European Society of Endocrinology reviewed and commented on preliminary drafts of these guidelines.

**Conclusions:** Continuous subcutaneous insulin infusion and continuous glucose monitoring have an important role in the treatment of diabetes. Data from randomized controlled trials are limited on the use of medical devices, but existing studies support the use of diabetes technology for a wide variety of indications. This guideline presents a review of the literature and practice recommendations for appropriate device use.

## Summary of Recommendations:

### 1. Insulin Pump Therapy Without Sensor Augmentation

1.1 We recommend continuous subcutaneous insulin infusion (CSII) over analog-based basal-bolus multiple daily injections (MDI) in patients with type 1 diabetes mellitus (T1DM) who have not achieved their A1C goal, as long as the patient and caregivers are willing and able to use the device. (1 ⊕⊕⊕○)

1.2 We recommend CSII over analog-based basal-bolus MDI in patients with T1DM who have achieved their A1C goal but continue to experience severe hypoglycemia or high glucose variability, as long as the patient and caregivers are willing and able to use the device. (1 ⊕⊕○○)

1.3 We suggest CSII in patients with T1DM who require increased insulin delivery flexibility or improved satisfaction and are capable of using the device. (2 ⊕⊕OO)

## 2. Insulin Pump Therapy in Type 2 Diabetes

2.1 We suggest CSII with good adherence to monitoring and dosing in patients with type 2 diabetes mellitus (T2DM) who have poor glycemic control in spite of intensive insulin therapy, oral agents, other injectable therapy, and lifestyle modifications. (2 ⊕⊕OO)

## 3. Insulin Pump Use in the Hospital

3.1 We suggest that clinicians continue CSII in patients admitted to the hospital with either type of diabetes, if the institution has clear protocols for evaluating patients as suitable candidates and appropriate monitoring and safety procedures. (2 ⊕⊕OO)

## 4. Selection of Candidates for Insulin Pump Therapy

4.1 We recommend that before prescribing CSII, clinicians perform a structured assessment of a patient's mental and psychological status, prior adherence with diabetes self-care measures, willingness and interest in trying the device, and availability for the required follow-up visits. (1 ⊕⊕OO)

## 5. Use of Bolus Calculators in Insulin Pump Therapy

5.1 We suggest encouraging patients to use appropriately adjusted embedded bolus calculators in CSII and have appropriate education regarding their use and limitations. (2 ⊕⊕OO)

## 6. Real-Time Continuous Glucose Monitors in Adult Outpatients

6.1 We recommend real-time continuous glucose monitoring (RT-CGM) devices for adult patients with T1DM who have A1c levels above target and who are willing and able to use these devices on a nearly daily basis. (1 ⊕⊕⊕⊕)

6.2 We recommend RT-CGM devices for adult patients with well-controlled T1DM who are willing and able to use these devices on a nearly daily basis. (1 ⊕⊕⊕⊕)

## Use of Continuous Glucose Monitoring in Adults with Type 2 Diabetes

6.3 We suggest short-term, intermittent RT-CGM use in adult patients with T2DM (not on prandial insulin) who have A1c levels  $\geq 7\%$  and are willing and able to use the device. (2 ⊕⊕OO)

## Education and Training on the use of Continuous Subcutaneous Insulin Infusion and Continuous Glucose Monitoring

6.4 We suggest that adults with T1DM and T2DM who use CSII and continuous glucose monitoring (CGM) receive education, training, and ongoing support to help achieve and maintain individualized glycemic goals. (Ungraded Good Practice Statement)

## Method: of Development of Evidence-Based Clinical Practice Guidelines

The Clinical Guidelines Subcommittee (CGS) of the Endocrine Society deemed diabetes technology a priority area in need of practice guidelines and appointed a Task Force to formulate evidence-based recommendations. The Task Force followed the approach recommended by the Grading of Recommendations, Assessment, Development, and Evaluation group, an international group with expertise in the development and implementation of evidence-based guidelines (1). A detailed description of the grading scheme has been published elsewhere (2). The Task Force used the best available research evidence to develop the recommendations. The Task Force also used consistent language and graphical descriptions of both the strength of a recommendation and the quality of evidence. In terms of the strength of the recommendation, strong recommendations use the phrase "we recommend" and the number 1, and weak recommendations use the phrase "we suggest" and the number 2. Cross-filled circles indicate the quality of the evidence, such that ⊕OOO denotes very low quality evidence; ⊕⊕OO, low quality; ⊕⊕⊕O, moderate quality; and ⊕⊕⊕⊕, high quality. The Task Force has confidence that persons who receive care according to the strong recommendations will derive, on average, more good than harm. Weak recommendations require more careful consideration of the person's circumstances, values, and preferences to determine the best course of action. Linked to each recommendation is a description of the evidence and the values that the

Task Force considered in making the recommendation; in some instances, there are remarks, a section in which the Task Force offers technical suggestions for testing conditions, dosing, and monitoring. These technical comments reflect the best available evidence applied to a typical person being treated. Often this evidence comes from the unsystematic observations of the Task Force and their values and preferences; therefore, one should consider these remarks as suggestions.

In this guideline, the Task Force made several statements to emphasize the importance of shared decision-making, general preventive care measures, and basic principles of diabetes technology. They labeled these "Ungraded Good Practice Statement". Direct evidence for these statements was either unavailable or not systematically appraised and considered out of the scope of this guideline. The intention of these statements is to draw attention and remind providers of these principles; one should not consider these statements as graded recommendations (3).

The Endocrine Society maintains a rigorous conflict-of-interest review process for developing clinical practice guidelines. All Task Force members must declare any potential conflicts of interest by completing a conflict-of-interest form. The CGS reviews all conflicts of interest before the Society's Council approves the members to participate on the Task Force and periodically during the development of the guideline. All others participating in the guideline's development must also disclose any conflicts of interest in the matter under study, and a majority of these participants must be without any conflicts of interest. The CGS and the Task Force have reviewed all disclosures for this guideline and resolved or managed all identified conflicts of interest.

Conflicts of interest are defined as remuneration in any amount from commercial interests; grants; research support; consulting fees; salary; ownership interests (e.g., stocks and stock options [excluding diversified mutual funds]); honoraria and other payments for participation in speakers' bureaus, advisory boards, or boards of directors; and all other financial benefits. Completed forms are available through the Endocrine Society office.

The Endocrine Society provided all funding for this guideline; the Task Force received no funding or remuneration from commercial or other entities.

#### Commissioned Systematic Review:

The Task Force commissioned a systematic review and individual patient data meta-analysis (Benkhadra, Real Time Continuous Glucose Monitoring in Type 1 Diabetes: A Systematic Review and Individual Patient Data Meta-Analysis, *JCEM [In Press]*) of randomized controlled trials (RCTs) that enrolled individuals with T1DM and compared RT-CGM versus control groups. A two-step regression model was used to pool individual patient data, which came from trial lists and device manufacturers. Pooled data from 11 RCTs suggested that the use of RT-CGM was associated with a significant reduction in HbA1c (-0.276; 95% confidence interval -0.465 to -0.087). The improvements in HbA1c were primarily seen in ages  $\geq 15$  years. There was no statistically significant difference in time spent in hypoglycemia or the number of hypoglycemic episodes, although these analyses were imprecise and warrant lower confidence. There was no difference between males and females.

**T**he goal of glucose management in all types of diabetes is to minimize and hopefully eliminate the acute and chronic complications associated with diabetes, such as the risks of microvascular complications and potentially (to a lesser degree) macrovascular complications and mor-

tality (4–9). All persons with T1DM require insulin, and persons with T2DM frequently need insulin for adequate glucose control. Patients requiring intensive insulin therapy take insulin as needed, adjusting both basal and prandial doses to reach specific glucose goals (10). However,

ISSN Print 0021-972X ISSN Online 1945-7197

Printed in USA

Copyright © 2016 by the Endocrine Society

Received June 30, 2016. Accepted August 12, 2016.

Abbreviations: CGM, continuous glucose monitoring; CSII, continuous subcutaneous insulin infusion; CGS, Clinical Guidelines Subcommittee; T1DM, type 1 diabetes mellitus; T2DM, type 2 diabetes mellitus; U.S., United States; FDA, Food and Drug Administration; MDI, multiple daily injections; SMBG, self-monitoring blood glucose; RAA, rapid-acting analog; RCT, randomized controlled trials; JDRF, Juvenile Diabetes Research Foundation; RT-CGM, real-time continuous glucose monitoring; SIP, sensor-integrated pump; SAP, sensor-augmented pump

intensification increases the risk of hypoglycemia, which is associated with both morbidity and mortality. Advances in the pharmacokinetics and pharmacodynamics of insulin products and in the methods of insulin delivery and glucose monitoring are geared towards improving glucose control, minimizing hypoglycemia, and improving quality of life (QOL). Two such advances include CSII and CGM. And while these and other new technologies hold the potential for enhancing outcomes and improving QOL for people with diabetes, reliable data on the efficacy of new advances is lacking. This is due in part because the United States (U.S.) Food and Drug Administration (FDA) regulates medical devices less rigorously than they regulate pharmaceuticals, which lessens the demand for large clinical trials comparing one device to another. Of the studies that do exist, many include devices that are outdated or obsolete at the time of publication and thus do not reflect current technologies. This guideline attempts to assess all available data on existing and emerging technologies and procedures for improving glucose control for patients with diabetes.

Important to note, success of these devices and technologies is directly linked to the level to which people are educated, capable, and willing to use them. Patients on CSII therapy, as well as MDI of insulin, can have inadequate glucose control. Therefore, in every stage of this guideline we stress the need for patient and practitioner education and training in conjunction with clearly defined processes for patient follow-up, analysis of device data, and access to care (should a problem arise). Finally, human factors relating to the use of these technologies are as important as other considerations in device development; although in many cases true effectiveness data are lacking (11).

Clinicians began developing CSII therapy as early as 1963, and it gained increasing general acceptance in the 1980s. This was in part due to the Diabetes Control and Complications Trial, which compared CSII with MDI in patients with T1DM. The trial lasted from 1982 to 1993 and placed roughly 40% of patients on CSII and the remaining patients on MDI (12). Results showed that patients treated with CSII had a slightly lower A1c mean relative to patients treated with MDI. These results, however, do not necessarily apply to today's diabetes patients, as the last 30 years has seen many technological improvements in insulin products and pump technologies, and the use of insulin pumps has grown dramatically (primarily for treating T1DM). In spite of these improvements, comprehensive reliable data are lacking on the number of people with T1DM and the number of people who use insulin pumps who have T1DM. Likewise, data are lacking that demonstrates a clear superiority of CSII over MDI. This is

due to a general lack of studies and confounded by the ongoing development of newer insulin analogues and advanced glucose monitoring and insulin delivery technologies, making comparisons between studies complex. However, diabetes specialists have confidence in pump therapy, despite its greater expense. The T1D Exchange, a registry of diabetes specialty clinics in the U.S., reported in 2012 that 56% of the adult T1DM patients in these clinics were using CSII and had lower HbA1c levels than those using MDI (13).

Early pumps were heavy, crude syringe pumps with suboptimal quality control (QC), inadequate battery power, and limited dosing flexibility that used infusion sets with rigid needles. The modern age of CSII use in the clinical setting began in the 1980s with new manufacturers entering the market and significant technological advances. Today, there are five FDA-approved pumps available in the U.S. providing numerous features to improve accuracy, safety, dosing decisions, convenience, and overall usability (14). This guideline only reviewed pumps that allow for multiple basal rates and use bolus dose calculators. We feel there is insufficient research evidence and experience to include single-basal-rate pumps or bolus-only pumps (Calibra Finesse ©) in this analysis. Finally, these guidelines are meant for individuals using rapid-acting analog (RAA) insulin in their pumps (although we do mention U-500 insulin).

There are few professional organization recommendations or guidelines related to CSII (15–17). The latest standards of care from the American Diabetes Association simply say that clinicians should treat most people with T1DM with MDI or CSII (10).

CGM is a more recently developed device for managing diabetes. For years, clinicians have used standard capillary blood glucose measurements (self-monitoring blood glucose [SMBG]) to guide therapy; however, these discrete values offer only a limited perspective on the constant daily changes in blood glucose levels, do not provide alarms that indicate when blood glucose levels are above or below various thresholds, and do not indicate trends in blood glucose levels. Current models of CGM measure the glucose concentration in the interstitial fluid, and devices are evolving steadily in terms of accuracy and ease of use. We reviewed the data regarding current CGM devices, giving specific consideration to more recent device technology. We excluded devices designed for intensive care units and the Abbott Libre © system, which, although unique in terms of its claim as strip replacement without the need for calibration, does not provide true CGM in terms of providing alerts for high and low blood glucose levels.



## 1. Insulin Pump Therapy Without Sensor Augmentation

1.1 We recommend CSII over analog-based basal-bolus MDI in patients with T1DM who have not achieved their A1C goal, as long as the patient and caregivers are willing and able to use the device. (1 ⊕⊕⊕O)

### Evidence

The alternative therapy for T1DM is basal-bolus therapy with a RAA insulin and a basal analog insulin (10). Unfortunately, many of the comparative effectiveness trials for pumps took place before these modern analogs (which are associated with less hypoglycemia) were available (18). Some studies compared CSII using RAA vs MDI using nonanalog or a mixture of insulin types; however, this is not directly applicable to current MDI regimens, which primarily use analog insulin. Continual advancements in pump design have also complicated analysis. The user interface, size and shape, and bolus calculators on pumps are evolving, as is data communication (eg, via CGM, meters, Internet). The added benefit of advancements in pump design would be obscured in a meta-analysis of multiple studies that also included outdated pumps. Finally, double-blind trials are not possible since individuals would be aware of their assigned treatment.

Given these limitations, however, a couple of meta-analyses have reviewed the small number of comparative effectiveness RCTs of therapies for T1DM in adults (18–23). The 2010 Cochrane Review of CSII in T1DM found a statistically significant reduction in A1c of 0.3% in those adults treated with CSII (19). However, that review included many studies completed prior to 2000 with pumps that used regular insulin. Due to the changes in available insulin and the technological changes in pumps, the most relevant systematic reviews are those evaluating studies reported over the last 15 years with pumps that used RAAs. The most recent meta-analysis by Yeh et al only included studies that used RAA and only found four qualifying trials (20). They again found a significant benefit to CSII with a reduction in A1c of 0.3% compared to MDI. However, the results were significantly influenced by the largest study that found a reduction in A1c of 0.84% with CSII; however, this was also the study with the highest randomization A1c (9.3% mean) (24). This 16-week study was intended to be a 32-week crossover trial, but they only reported the first phase due to a high dropout rate at the time of crossover. In spite of this significant flaw, this study was unique in that it included a 14-week “qualification phase” to randomize only patients who completed at least 70% of the recommended glucose monitoring. There is no evidence that other studies took measures to identify patients for inclusion who were similarly

motivated. Only one of the reported studies was of more than 16 weeks in duration. In a small 9-month long study, Tsui et al reported that baseline A1c was approximately 0.5% higher in the MDI group, although they said it represented a nonsignificant difference (25). None of the studies in the Yeh analysis used a bolus dose calculator, although one of the studies included a pump that did have that feature available (26). This study was the only one that used a modern basal insulin (glargine).

A 5-week cross-over design study using CGM reported that the area under the curve (AUC) for glucose levels over 140 mg/dL was reduced by 40% in T1DM patients on CSII of insulin aspart vs MDI of insulin aspart/insulin glargine (27). Fructosamine reduction was also highly significant with CSII vs MDI.

Observational studies performed at clinics with significant insulin pump experience and carefully applied protocols of patient selection and education suggest that glucose control may be better in these clinics vs research settings where patient inclusion criteria are variable and rarely optimized for ideal candidates. For example, a recent observational study of 200 adults with T1DM transitioned to CSII from MDI demonstrated that in these poorly controlled patients CSII reduced HbA1c by a mean of > 1.0% and maintained a significant improvement for the average 6 years of follow-up (28). RCTs do not consistently utilize (and meta-analyses do not typically analyze) the selection criterion, educational approach, or ongoing evaluation and support that experienced clinicians and clinics implement. We describe in detail what defines a reasonable pump candidate in section 4.

Despite the limitations of the available literature, there is relatively consistent evidence that current CSII is likely to improve glucose control in motivated patients with inadequate glucose control who are appropriately educated and supported. Because this area of insulin treatment technology is progressing in the direction of sensor augmentation, it is unlikely that we will see meaningful studies evaluating the isolated benefit of CSII that will advance the present body of evidence.

1.2 We recommend CSII over analog-based basal-bolus MDI in patients with T1DM who have achieved their A1C goal but continue to experience severe hypoglycemia or high glucose variability, as long as the patient and caregivers are willing and able to use the device. (1 ⊕⊕OO)

### Evidence

Data on the reduction of hypoglycemia are mixed and difficult to analyze with systematic reviews due to variable study definitions of hypoglycemia and variable patterns of monitoring to detect overall or nocturnal hypoglycemia (19, 20). On balance, there is no good evidence that CSII

reduces overall hypoglycemia in patients with T1DM. Most comparative trials of CSII vs MDI have excluded patients experiencing severe hypoglycemia (usually defined as requiring assistance) in the months before entry into the study (19, 22). In studies where these patients are specified, meta-analysis shows that severe hypoglycemia is significantly reduced on CSII compared to MDI, though these studies used insulin regimens that are outdated (29).

There is also evidence that CSII is associated with reduced glucose variability (30–33). This is important because variability is one of the primary indications for CSII in some clinics (34). Furthermore, higher variability is often associated with more hypoglycemia and there is an unproven concern that variability may have an independent effect on complications (35).

1.3 We suggest CSII in patients with T1DM who require increased insulin delivery flexibility or improved satisfaction and are capable of using the device. (2 ⊕⊕○○)

## Evidence

As noted above, there is no evidence from systematic reviews that indicates inferior glucose control with CSII relative to MDI (19, 21, 36). Therefore, expanded CSII use is limited by higher cost and marginal benefit, unless there are other advantages beyond glucose control. The flexibility provided by CSII with RAA insulin could be an advantage for those who exercise and potentially those with gastroparesis, as the basal delivery dose and pattern can be modified (37–41). A case series of patients with gastroparesis who converted from MDI to CSII showed fewer hospitalizations, improved HbA1c, and less glucose variability (41). However, there are no RCTs supporting this potential benefit. Many studies have demonstrated improved QOL or improved patient satisfaction with CSII therapy relative to MDI therapy, some of which may be due to improvements in glycemic control. However, a patient's attitude towards different technologies and methods of implementation, while significant, is not easily defined.

Unfortunately, various studies used different measurement tools or reported on different specific categorical findings (42). Therefore, while almost all positive findings are in favor of CSII, the inconsistencies provide limited clarity in meta-analysis (20, 24–26, 32, 43, 44).

## 2. Insulin Pump Therapy in type 2 Diabetes

2.1 We suggest CSII with good adherence to monitoring and dosing in patients with T2DM who have poor glycemic control in spite of intensive insulin therapy, oral agents, other injectable therapy, and lifestyle modifications. (2 ⊕⊕○○)

## Evidence

Although several uncontrolled studies have reported that T2DM patients treated with CSII have improved glucose control and patient-reported outcomes, RCTs have shown mixed results and subsequent meta-analyses have failed to show significant reductions of A1c or reductions in hypoglycemia for T2DM patients on CSII (20, 22, 45–51). Additionally, the older studies did not compare CSII to MDI using insulin analogues. A recent RCT of 331 T2DM patients on MDI randomized to CSII or MDI using analog insulin reported a statistically superior reduction in A1c of 1.1% from the baseline mean of 9.0% in the CSII group; the MDI group experienced a 0.4% reduction from the same baseline A1c (52). In contrast to other studies, this study only enrolled patients with an A1c level between 8.0–10.0% who demonstrated adherence to monitoring during a 2-month run-in period. Hypoglycemia rates were not different between the two groups. Based on cost and limited generalizable evidence, CSII does not represent standard of care for the routine patient with T2DM who requires insulin therapy. However, this study indicates that properly chosen patients on advanced insulin therapy may benefit from CSII. New pharmacologic treatments, such as sodium-glucose cotransporter 2 inhibitors and glucagon-like peptide-1 agonists, can delay the progression to intensive basal bolus therapy and further reduce the need for CSII in T2DM (53).

The above statements all pertain to the use of U-100 analog insulin in pumps. Nonrandomized retrospective studies that moved patients from U-100 insulin MDI to U-500 insulin by CSII have demonstrated a significant reduction in A1c (54–57). However, there have been no RCTs that have compared CSII with MDI using U-500 regular insulin. The only comparison in the literature is based on a loosely designed meta-analysis that compared U-500 by MDI vs CSII between studies, suggesting equal efficacy but less weight gain and less increase in insulin dose with CSII (56). Because MDI of U-500 insulin is quite effective in highly insulin-resistant patients, we will need to clearly demonstrate that CSII of U-500 insulin is more effective to justify the added expense (58, 59).

Until we have adequate reliable data supporting the benefit of CSII vs MDI of U-500 insulin, we suggest limiting this approach to patients whose total daily insulin dose is greater than 200 U/d in spite of using all alternative modalities for treating T2DM. Furthermore, a provider who is knowledgeable about CSII of U-500 should supervise its use.

## 3. Insulin Pump Use In The Hospital

3.1 We suggest that clinicians continue CSII in patients admitted to the hospital with either type of diabetes, if the

institution has clear protocols for evaluating patients as suitable candidates and appropriate monitoring and safety procedures. (2 ⊕⊕OO)

### Evidence

No RCTs have evaluated the relative benefit of continuing CSII vs transitioning to intravenous (IV) insulin infusions or MDI therapy when patients on ambulatory pump therapy are hospitalized for acute medical illness or surgery. However, studies have reported that hospitals that have well-developed criterion and processes for diabetes care can deliver safe and effective in-patient CSII therapy (60–65).

Both the American Diabetes Association and the American Association of Clinical Endocrinologists support these findings, assuming the patient has the physical and mental capacity for continued CSII use (15, 66). There are factors (eg, medical illness, comedications, the degree of acute insulin resistance, mental status changes) that could determine if CSII therapy is suitable for a hospitalized patient; however, no studies have systematically investigated these factors. Therefore, institutions that are unable to guarantee appropriate evaluation and support may decide against CSII use for inpatients (for reasons such as concerns for medical liability). It is often recommended that this evaluation be completed by an endocrinologist experienced in insulin pump therapy. No specific evidence exists in this regard and it is likely that any provider with in depth pump knowledge and experience could appropriately guide this process. If CSII is discontinued, appropriate transition to basal bolus therapy is imperative.

CSII can generally be continued in patients undergoing same-day or outpatient surgery, which may involve fasting and/or conscious sedation. However, clinicians need to know the patient is on an insulin pump and provide him/her with recommendations on how to prepare for surgery with regards to pump settings. If clinicians are not comfortable with continuing CSII therapy, they should give insulin as an injection.

### 4. Selection of Candidates for Insulin Pump Therapy

4.1 We recommend that before prescribing CSII, clinicians perform a structured assessment of a patient's mental and psychological status, prior adherence with diabetes self-care measures, willingness and interest in trying the device, and availability for the required follow-up visits. (1 ⊕⊕OO)

### Evidence

There are few studies and no RCTs or systematic reviews that have specifically examined or identified which

factors predict successful CSII use (15, 19–22, 29, 67), and some patients do well with CSII who don't exhibit characteristics that are considered favorable for success. Even less evidence is available to determine what knowledge, expertise, and resources clinicians need to teach and support CSII use. Of the studies that did examine factors for successful CSII use, there is relatively consistent evidence that higher HbA1c levels at baseline are associated with greater A1c reduction on CSII (28, 29, 68–70). However, Orr et al reported that an A1c level over 10% was associated with poor outcomes with CSII (28). Interestingly, Nixon et al recently reported that among those placed on CSII for elevated A1c, about 12% show no benefit at any time after transitioning from MDI, and 57% showed early A1c reduction with later deterioration (71). In regards to other factors, the study by Orr et al found that mental illness and a history of missed appointments predicted worse outcomes with CSII (28). Likewise, Grant et al reported that significant anxiety or depression was associated with poorer outcomes with CSII (72). Data are lacking on the relationship between the frequency of SMBG and success with CSII. However, one study that evaluated clinic experience did report an association between the frequency of SMBG and success with CSII (73). Among adults, age does not appear to be a major determinant (74).

In summary, there is modest evidence that certain criteria are relevant in choosing patients for CSII therapy. Clinical experience suggests that adult CSII candidates should receive a thorough evaluation prior to initiating therapy to assess a wide-range of diabetes self-care behaviors. Factors to consider are: baseline A1c, patient history of adherence to follow-up appointments and provider-recommended glucose monitoring, a realistic recognition by the patient of the limitations of CSII, evidence of a significant psychological disease that is likely to impact adherence, and confirmed availability for follow-up in the immediate postpump initiation period. It is also recommended that prior to initiating CSII, clinicians assess and address general gaps in patient diabetes knowledge or self-care behaviors, including carbohydrate counting and sick-day rules.

### 5. Use of Bolus Calculators in Insulin Pump Therapy

5.1 We suggest encouraging patients to use appropriately adjusted embedded bolus calculators in CSII and have appropriate education regarding their use and limitations. (2 ⊕⊕OO)

### Evidence

Many intensively managed individuals rely on estimations to calculate prandial and correction insulin (75). As



time passes, there may be a tendency to approximate the prandial bolus based on the “usual” or “typical” carbohydrate content of the meals with variable success (76). In 2002, the Deltec Cozmo © pump first introduced technology that helps patients calculate bolus insulin (carbohydrate to insulin ratios as well as correction ratios in the event of hyperglycemia) as part of CSII therapy. Other companies soon introduced their own calculators for this purpose, hoping to reduce variability and patient errors (77). Studies evaluating the effect of a dose calculator have found some beneficial outcomes, such as reduced mean glucose, reduced need for glucose treatment of hypoglycemia, decreased frequency of correction boluses (reduced stacking), and decreased postprandial glucose levels (78–80). However, findings have not been consistent, and no large RCTs have specifically demonstrated the benefit of bolus calculators.

The calculator feature was unique to CSII, but this capability is now available for those on MDI, where it has reduced A1c by at least 0.5% and improved patient satisfaction but not hypoglycemia (81–83). In order to facilitate the transition from one therapeutic approach to another and to help achieve glycemic targets, we suggest that prior to initiating CSII therapy, clinicians should determine individualized prandial and correction insulin dosing algorithms (insulin-to-carbohydrate ratio and insulin sensitivity factors), blood glucose targets, and active insulin time (84). We do not recommend the general use of insulin calculators unrelated to CSII, such as smart phone apps that are not FDA approved. The FDA has only approved the meter bolus calculator mentioned above (69).

The current approach to prandial insulin dose calculation (based on carbohydrate-to-insulin ratios) assumes that carbohydrates are the only macronutrient that impact postprandial glucose control in T1DM. However, as summarized in a recent systematic review of the literature (85), high-fat/protein meals require more insulin than lower-fat/protein meals with identical carbohydrate content in order to optimize postprandial glucose control (86, 87). However, Bell et al (85) reported that available studies have significant methodological differences (88–90), and reliable definitive data are lacking on the optimum split and duration of advanced pump boluses needed to adjust glucose levels associated with high fat and protein content meals.

## 6. Real-Time Continuous Glucose Monitors in Adult Outpatients

6.1 We recommend RT-CGM devices for adult patients with T1DM who have A1c levels above target and who are willing and able to use these devices on a nearly daily basis. (1 ⊕⊕⊕⊕)

### Evidence

The Juvenile Diabetes Research Foundation (JDRF) study (91), the Guard Control Study (92), and O’Connell et al (93) demonstrated that adults with HbA1c levels  $\geq 7.0\%$  had a greater reduction in A1c using RT-CGM than with intermittent SMBG (0.5%, 0.6%, and 0.43%, respectively). Furthermore, unlike findings with SMBG, the improvement in A1c with CGM is not accompanied by an increase in biochemical hypoglycemia (92, 94). The improvement in A1c in the CGM subjects in the 6-month JDRF trial was sustained during the 6-month observational period that followed completion of the trial (95). This ongoing benefit occurred despite a reduction in office-visit frequency during this observational period to levels similar to routine care ( $2.7 \pm 1.2$  visits over 6 months). Furthermore, the incidence rate of severe hypoglycemia declined from 20.5 events per 100 patient-years during the initial 6-month RCT to 12.1 events per 100 patient-years during the 6-month observational follow-up period.

To date, CGM trials have enrolled patients using continuous CSII pumps and MDI, and no studies have evaluated the added benefit of starting CGM in adult patients using MDI therapy. In the JDRF trial (91), the patients using CSII and MDI had a similar reduction in A1c; however, because MDI users comprised only 20% of the total study population, the improvement in this subgroup did not reach statistical significance. The ongoing Multiple Daily Injections and Continuous Glucose Monitoring in Diabetes study (96) examining RT-CGM in T1DM and T2DM adults using MDI should provide conclusive data regarding the benefit of this technology in these individuals.

6.2 We recommend RT-CGM devices for adult patients with well-controlled T1DM who are willing and able to use these devices on a nearly daily basis. (1 ⊕⊕⊕⊕)

### Evidence

The JDRF study demonstrated that when compared with standard blood glucose monitoring, T1DM patients with A1c levels less than 7.0% that use RT-CGM can reduce the frequency of biochemical hypoglycemia (which they defined as a blood glucose level below 70 mg/dL) and maintain A1c levels less than 7.0% over a 6-month study period. Of the 129 enrolled subjects, 62 (48%) were younger than 25, and 67 (52%) were more than 25 years of age. The median time per day with a glucose level of  $\leq 70$  mg/dL was less in the RT-CGM group than in the control group; however, the difference was not statistically significant. In this study, almost all the other analyses (including the time per day  $\leq 60$  mg/dL, time per day between 71 and 180 mg/dL, and combined outcomes involving A1c coupled with hypoglycemia) statistically favored the RT-



CGM group compared with the control group. Treatment effects were generally similar across age groups (95). For RT-CGM users who were 25 years and older, the incidence rate of severe hypoglycemia was 21.8 events per 100 person-years during the 6-month RCT and 7.1 events per 100 person-years during the 6 months of CGM use following the trial. For those in this group whose A1c levels were below 7.0%, the incidence rate of severe hypoglycemia was 23.6 events per 100 person-years during the 6-month RCT and 0 per 100 patient-years during the 6 months of CGM use following the trial (97). This evidence of improvements in glycemic control over the long-term points to the role of the user's skills and knowledge of new CGM technology, and this may partly account for the failure of other RCTs (that enrolled individuals with poorer glycemic control) to demonstrate a reduction in severe hypoglycemia (98, 99). In a multicenter European/Israeli RCT that included adults with T1DM whose HbA1c levels were less than 7.5%, a post hoc per protocol analysis demonstrated that time spent in hypoglycemia below 63 mg/dl was reduced by 50% ( $P = .02$ ) in the adults (100).

### Use of Continuous Glucose Monitoring in Adults with type 2 Diabetes

6.3 We suggest short-term, intermittent RT-CGM use in adult patients with T2DM (not on prandial insulin) who have A1c levels  $\geq 7\%$  and are willing and able to use the device. (2 ⊕⊕○○)

#### Evidence

This recommendation is based on data from one well-performed large-scale RCT performed by Vigersky et al (101, 102) and one other smaller study (103) showing similar outcomes. The Vigersky trial involved 100 adults with T2DM on therapies including diet and exercise alone and various combinations of other antihyperglycemic medications, including basal but not prandial insulin. Results showed that intermittent RT-CGM use for 12 weeks (4 cycles of 2 weeks use/1 week off) resulted in significant improvements in A1c, sustained during a 40-week follow-up period (the changes in A1c [mean  $\pm$  SEM] at 12 weeks and 52 weeks were  $1.0 \pm 1.1\%$  and  $0.8 \pm 1.5\%$  in the RT-CGM group vs.  $0.5 \pm 0.8\%$  and  $0.2 \pm 1.3\%$  in the SBGM control group,  $P = .04$  for control vs. RT-CGM group). This improvement in the RT-CGM group occurred without a greater intensification of medication, as compared the control group, indicating that it probably reflected changes in self-care prompted by CGM use. Yoo et al (103) randomized patients (with T2DM and A1c levels of 8%–10% on oral agents or insulin) to RT-CGM use for 3 days per month for 12 weeks vs SBGM 6 times

per week for 12 weeks in the control group. Both the RT-CGM and SBGM control groups had significant reductions in A1c at 12-week follow-up (RT-CGM:  $9.1\% \pm 1.0\%$  to  $8.0 \pm 1.2\%$ ,  $P < .001$ ; SBGM:  $8.7 \pm 0.7\%$  to  $8.3 \pm 1.1\%$ ,  $P = .01$ ) with a significant difference in the improvement between the two groups ( $P = .004$ ). The relative difference in improvements in A1c between the RT-CGM and control groups may have been exaggerated by the fact that the control group only performed SBGM 6.1 times per week in the trial, and this may limit the relevance of the trial findings, especially to patients who perform SBGM multiple times per day.

We need consistent data from additional well-performed RCTs in patients with different health literacy and socio-demographic characteristics (compared to patients enrolled in the above 2 trials) to confirm that these findings are generalizable to the broader T2DM population. At present, no data exist for RT-CGM use in patients with T2DM on prandial insulin. The ongoing Daily Injections and Continuous Glucose Monitoring in Diabetes study (96) should provide conclusive data regarding the potential benefits of this technology in this T2DM population.

We excluded several studies on CGM in T2DM that were incorporated in a previous meta-analysis (104), because they employed older technology (105) and involved blinded (instead of real-time) CGM use (106). Retrospective data analysis from blinded CGM devices has a role in the care of patients with both T1DM and T2DM, especially the elderly population who, because of restrictions on Medicare coverage, do not have access to RT-CGM. This older population has a relatively high incidence of hypoglycemia. Munshi et al demonstrated that blinded CGM is helpful in therapeutic decision-making in this vulnerable age group, and it also helps detect hypoglycemia that is otherwise unrecognized with intermittent capillary blood glucose monitoring (107).

### Education and Training on the use of Continuous Subcutaneous Insulin Infusion and Continuous Glucose Monitoring

6.4 We suggest that adults with T1DM and T2DM who use CSII and CGM receive education, training, and ongoing support to help achieve and maintain individualized glycemic goals. (Ungraded Good Practice Statement)

#### Evidence

There are few high-quality comparative studies regarding the effectiveness of educational components and strategies for CSII use. As a result, there are limited data about how to best train individuals on how to use CSII to optimize glycemic control. It should also be noted that primary care providers manage a good number of patients using

CSII, and in this setting insulin pump company employees or contracted consultants provide initial training. We found no studies that compared this form of education and training to that delivered in an endocrinology practice by diabetes educators using a formal CSII education curriculum. No RCTs have evaluated the effectiveness of educational strategies and protocols for transitioning patients from MDI to CSII with stand-alone RT-CGM or sensor-integrated pump therapy (SIP) (where the CGM is integrated with the insulin pump but does not necessarily control insulin delivery, as occurs with sensor-augmented pump therapy [SAP]). It is important to be aware of insurance plan requirements with regards to participation in a comprehensive diabetes program. Some require participation in these programs within 6 months prior to either a new pump application or CGM application; this is also true for patients seeking an upgraded device.

### Continuous Subcutaneous Insulin Infusion

A 2011 systematic review that included five descriptive studies of individuals with T1DM over the age of 16 recommended that adults with T1DM starting CSII or already utilizing CSII receive comprehensive advice, education, and training (84). The authors concluded, however, that it was difficult to draw strong conclusions about the effectiveness of educational components and strategies due to the lack of high-quality comparative studies. They concluded that no educational method was significantly

more effective than any other method. One descriptive study in 250 participants with T1DM found a highly significant impact on A1c levels ( $P < .0001$ ) and a reduction in hypoglycemia ( $P < .001$ ) over a 12-month period following a 7-day teaching and training program that included instructions regarding managing daily living, using CSII, adjusting insulin doses, preventing and managing hypoglycemia and diabetic ketoacidosis, and understanding the role of SMBG (108). Consensus statements and guidelines from the American Association of Clinical Endocrinologists and the American Association of Diabetes Educators also recommend education and training on technical aspects of pump operation, day-to-day self-care management, and how to handle emergency situations (15, 67). A pilot study ( $n = 30$ ) that adapted the 5-day Dose Adjustment for Normal Eating curriculum for people with T1DM reported that patients were receptive to starting CSII use, and CSII reduced HbA1c by 0.5% at 6 months (109, 110). Unfortunately even less research has been conducted on CSII education and training for individuals with T2DM (49, 50, 111). Table 1 outlines education and training for CSII therapy.

### Sensor-Integrated Pump Therapy or Continuous Subcutaneous Insulin Infusion with Stand-Alone Continuous Glucose Monitoring

To be successful with SIP or CSII, stand-alone CGM patients need to understand how to use this technology as

**Table 1.** Continuous Subcutaneous Insulin Infusion—Considerations for Education and Training

<b>Patient:</b>	<b>Time Periods to Assess Patient Self-Care Behaviors and Knowledge</b>		
<b>Collaborate with HCP overseeing CSII use and/or the multi-disciplinary diabetes team by returning for follow-up.</b>	<b>Prior to Initiating CSII Assess</b>	<b>Annually Re-assess</b>	<b>When Upgrading to a New CSII Device Re-assess</b>
Participate in using data management resources to make adjustments to therapy and evaluate self-care behaviors.			
<b>Provider:</b>	<b>When Discontinuing CSII or Transitioning to MDI Re-assess</b>		
Provide education as indicated to address deficiencies or when upgrading to new technology.			
Assess CSII use and evaluate for the loss of ability to operate insulin pump due to: cognitive, physical, or age-related changes; changes in insurance coverage; or changes in healthcare-provider-managing CSII use.			
	Glucose monitoring via SMBG frequency, and/or CGM use defined by HCP to meet individualized glycemic goals	Glucose monitoring via SMBG frequency and/or CGM use defined by HCP to meet individualized glycemic goals	Glucose monitoring via SMBG frequency and/or CGM use defined by HCP to meet individualized glycemic goals
	Carbohydrate counting or another method of mealtime bolus determination	Basal settings via basal rate testing across different time periods, adjust as indicated	New insulin plan for MDI
	Ability to operate CSII and make setting changes due to factors such as dexterity, vision impairment, mental health, or cognition---independently or with assistance from a designated care provider	Bolus calculator settings, if using feature, adjust as indicated	
	Infusion site health and selection	Infusion sites and type of infusion set, adjust as indicated	
	DKA prevention and treatment	Ability to troubleshoot insulin pump malfunction	
	Hypoglycemia---prevention, detection, and treatment	DKA prevention and treatment	
	Emergency supplies	Hypoglycemia prevention, detection, and treatment	
	If using bolus calculator, assess these settings:	Emergency supplies	
	Insulin-to-carbohydrate ratio	Back-up plan for use of injected insulin should pump fail	
	Insulin sensitivity factor		
	Glucose targets		
	Active insulin time		

Abbreviations: MDI, multiple daily injections. DKA, diabetic ketoacidosis; CGM, continuous glucose monitoring; SMBG, self-monitoring blood glucose; HCP, health care professional; CSII, continuous subcutaneous insulin infusion

[Derived from Powers *et al.* 2015 (113); Scheiner *et al.* 2009 (67); and Grunberger *et al.* 2014 (15)]

part of their daily diabetes self-management. Providing information about how these devices work prior to initiation, as well as supporting patients through education and training is key to safe and effective use of this technology. Unfortunately, research is lacking in this area as well. Although not designed to study the effectiveness of educational components associated with CGM use, two multicenter RCTs did provide detailed information on the education and training interventions provided in the treatment arms (91, 112). The JDRF CGM Study Group trial provided both the investigational and control groups with one-on-one training on RT-CGM or a glucose meter, as well as written instructions on how to make real-time adjustments of insulin doses and use computer software (91). Participants with SIP or CSII plus RT-CGM received additional instructions on modifying insulin doses by using treatment algorithms based on glucose trends. Another trial randomized participants with T1DM (who were naive to both CSII and RT-CGM) to either continue MDI or transition to SIP and provided a stepwise education protocol for introducing these different devices (112). The subjects in the SIP arm were introduced to CSII first and RT-CGM approximately 2 weeks later. The SIP arm also received technology-specific training over an additional 3-week period. The study used the patients' own data as a teaching tool to enhance experience-based training. In all subjects using SIP, A1c levels fell rapidly from baseline to 3 months (8.3+/-0.5% to 7.5% [absolute reduction of 0.8+/-0.8%,  $P < .001$ ]) and remained significantly lower throughout the study compared to those using MDI therapy (8.3+/-0.5% to 8.1% [absolute reduction of 0.2+/-0.9%,  $P < .001$ ]). In the absence of evidence from RCTs on training methods, clinical experience guides practice.

Table 2 outlines education and training for RT-CGM. Table 3 lists clinical and administrative resources that should be in place to support CSII and RT-CGM use.

### The Evolution of Continuous Subcutaneous Insulin Infusion and Continuous Glucose Monitoring Towards the Bionic Pancreas

Although SIP therapy is not new, there are no large RCTs comparing it directly to basal-bolus insulin injections using similar blood glucose monitoring strategies. However, there is an RCT involving 329 adult patients with T1DM that showed that SIP resulted in a 0.6% greater A1c reduction vs MDIs of analog insulin (112). The study did not include an arm with CSII alone or MDI with RT-CGM.

A new product, called SAP, has advanced SIP technology by adding a "low glucose suspend" function. A sensor reads glucose levels and adjusts the insulin pump accordingly, discontinuing insulin delivery when glucose reaches a programmed level. One study in adults with documented nocturnal hypoglycemia randomly assigned 247 patients to an SIP or an SAP that could stop insulin delivery for up to 2 hours (116). Overall nocturnal hypoglycemia (defined as 65 mg/dl or less for at least 20 minutes between 10 p.m. and 8 a.m.) was reduced by 31.8% without rebound hyperglycemia (1.5 +/- 1.0 vs. 2.2 +/- 1.3 per patient week,  $P < .001$ ); there were four episodes of severe hypoglycemia (defined as hypoglycemia resulting in coma or seizures or requiring medical assistance) all in the SIP-alone group. This RCT indicates that SAP can reduce nocturnal hypoglycemia in populations at high risk for hypoglycemia.

Current technologies, combined with faster insulin,

**Table 2.** Real-time Continuous Glucose Monitoring Technology—Considerations for Education and Training for Personal Use

Time Periods to Assess Patient Self-Care Behaviors and Knowledge		
Prior to Initiating RT-CGM Assess	When Initiating RT-CGM Assess	Annually and/or When Upgrading Technor Ongoing RT-CGM Use) Re-assess
<p><b>Patient:</b> Collaborate with HCP overseeing RT-CGM use and/or the multi-disciplinary diabetes team by returning for follow-up. Participate in use of data management resources to make adjustments to therapy and evaluate self-care behaviors.</p> <p><b>Provider:</b> Provide education as indicated to address deficiencies or when upgrading to new CGM technology. On an on-going basis assess CGM use and evaluate for the loss of ability to operate CGM system due to: cognitive, physical, or age-related changes; changes in insurance coverage; or changes in healthcare-provider-managing CSII use.</p>	<p>Knowledge of CGM system components---receiver, sensor, and transmitter</p> <p>Understanding of how CGM data differ from SMBG data</p> <p>Use of trend information based on changing glucose levels to adjust insulin doses</p> <p>Use of SMBG to calibrate CGM system</p> <p>Site selection and care</p> <p>Alarms (Check that alarms are set in a fashion to optimize patient benefit and minimize alarm fatigue. The initial focus is often on low alerts and falling BG alarms. High alerts can be turned off or set well above target at first if patient is consistently high and doesn't benefit from high alerts.)</p>	<p>Ability to make insulin adjustment based on trend information</p> <p>Use of SMBG to calibrate</p> <p>Sensor site health and care</p>

Abbreviations: MDI, multiple daily injections. DKA, diabetic ketoacidosis; CGM, continuous glucose monitoring; SMBG, self-monitoring blood glucose; HPC, health care professional; CSII, continuous subcutaneous insulin infusion

[Derived from Powers *et al.* 2015 (113); Evert *et al.* 2009 (114); and Gilliam *et al.* 2009 (115)]

**Table 3.** Suggested Health Care Provider Resources to Support the Safe and Effective Use of Continuous Subcutaneous Insulin Infusion and Real-time Continuous Glucose Monitoring

Clinicians should implement a formalized educational plan for the initiation and long-term support of CSII and RT-CGM that is established by the practice setting. Clinicians should communicate education plans to potential candidates and individuals new to the practice setting.

Clinics should have a designated multi-disciplinary diabetes team to evaluate potential candidates, initiate therapy and education, and to support long-term CSII and RT-CGM use that is part of the practice setting. Alternatively, a sole practitioner may also provide medical care and CSII education supported by other healthcare professionals or industry consultants as indicated. CSII and RT-CGM diabetes care team members could include the following health care providers: endocrinologist, ARNP, physician's assistant, primary care provider, certified diabetes educator, registered dietitian with expertise in diabetes medical nutrition therapy, and mental health/behavioral therapists.

The clinical support staff or patient should know how to download data from glucose meters, insulin pumps, and RT-CGMs, and they should download data prior to an appointment in order to review data via print format, online secure patient pump portal, or the patient's electronic health record.

Designated staff members (diabetes educators, medical assistants, or patient service representatives) should be available to help clinics obtain prior authorizations, schedule peer-to-peer reviews, submit supporting documents for determining coverage, answer insurance and distributor/third party vendor queries for additional documentation, and issue letters of support and/or appeal if therapy is declined.

Abbreviations: CGM, continuous glucose monitoring; CSII, continuous subcutaneous insulin infusion; APRN, advanced practice registered nurse; RT, real time

better sensor accuracy, and improved pumps that can adjust or discontinue insulin delivery to prevent hypoglycemia and excessive hyperglycemic exposure, make the promise of a "closed loop" seem reasonable. It is unclear at this time if a dual hormone system would be superior (117, 118). It is premature to grade the early data from studies on these new technologies, but the future for a commercial version of one or more of these devices is promising. In order to realize large-scale and long-term use of these devices, we need a better understanding of the catheter and site problems common with CSII and RT-CGM and appropriate cost-benefit analysis when compared to current therapies. A European Association for the Study of Diabetes and American Diabetes Association Diabetes Technology Working Group statement discusses suggestions for future research on the efficacy and safety of CSII (119). Finally, prior to the arrival of these new technologies, efforts have been ongoing to standardize the reporting of data from various devices so this information can be more easily interpreted and incorporated into electronic medical records (112). It is imperative that as devices proliferate, we consider the human factor, both in terms of how easily, safely, and effectively people with diabetes and their caregivers can use these devices and how easily and effectively caregivers can access data from these devices and provide feedback to patients in order to optimize care.

## Acknowledgments

Cosponsoring Associations: The American Association for Clinical Chemistry, the American Association of Diabetes Educators, and the European Society of Endocrinology. Special thanks are

extended to Drs. Anthony L. McCall and Timothy S. Bailey for careful review and thoughtful suggestions and Eric Vohr, medical writer, for excellent editorial assistance.

Address all correspondence and requests for reprints to: .

This work was supported by **Financial Disclosures of the Task Force\***: **Anne L. Peters, MD (chair)**—Financial or business/organizational interests: Medscape-Education (Advisory Board), ABIM-Board Examinations (Endocrine Subspecialty Board), American Diabetes Association (Speaker), U.S. FDA (Consultant) Lexicon (Advisory Board), NIH (Grantee to Institution), Helmsley Charitable Trust (Grantee to Institution), JDRF (Committee Member), Bigfoot Biomedical (Advisory Board), OptumRx (Ad Hoc Consultant), CVS/Caremark (Ad Hoc Consultant), Omada Health (Advisory Board); Significant financial interest or leadership position: Abbott Diabetes Care (Advisory Board), Astra Zeneca (Advisory Board), Biondi (Advisory Board), Lilly (Advisory Board/Speaker), BD (Advisory Board), Merck (Advisory Board), Janssen (Advisory Board/Speaker/Principal Investigator [Grantee]), Medtronic (Principal Investigator [Grantee]), Novo Nordisk (Advisory Board), Boehringer Ingelheim (Advisory Board). **Andrew J. Ahmann, MD, MS**—Financial or business/organizational interests: American Diabetes Association (Member/Regional Leadership Council), AACE (Member), Horizon CME (Consultant/Speaker for PriMed), Sanofi (Advisory Board/Research), Novo Nordisk (Consultant/Research Funding), Lilly (Consultant/Advisory Board), Janssen (Consultant/Advisory Board), Dexcom (Research Funding), Medtronic (Research Funding), Lexicon (Research Funding); Significant financial interest or leadership position: none declared. **Tadej Battelino, MD, PhD**—Financial or business/organizational interests: Medtronic (Advisory Board/Research Grant to Institution/Speakers Bureau), Novo Nordisk (Advisory Board/Research Grant to Institution/Speakers Bureau), Lilly (Advisory Board/Speakers Bureau), Sanofi (Advisory Board/Speakers Bureau), Bayer (Advisory Board), Abbott (Research Grant to Institution), Sandoz (Research Grant to Institution), Roche (Speakers Bureau); Significant financial interest or leadership position: DreaMed (Chief Medical Officer/Owner-



ship Interests). **Alison Evert, MS, RD, CDE**—Financial or business/organizational interests: none declared; Significant financial interest or leadership position: none declared. **Irl B. Hirsch, MD**—Financial or business/organizational interests: Wolters Kluwer (Editor), Abbott Diabetes Care (Consultant), Roche (Consultant), Intarcia (Consultant); Significant financial interest or leadership position: none declared. **M. Hassan Murad, MD, MPH\*\***—Financial or business/organizational interests: Mayo Clinic Evidence-based Practice Center; Significant financial interest or leadership position: none declared. **William E. Winter, MD**—Financial or business/organizational interests: none declared; Significant Financial interest or leadership position: none declared. **Howard Wolpert, MD**—Financial or business/organizational interests: Abbott Diabetes Care (Member/Advisory Board), BD (Member/Advisory Board), Bigfoot Biomedical (Member/Advisory Board), Dexcom (Consultant), Glooko (Member/Advisory Board), Insulet (Member/Advisory Board), Novo Nordisk (Member/Advisory Board), Zealand (Consultant); Significant financial interest or leadership position: Abbott Diabetes Care (Member/Advisory Board), BD (Member/Advisory Board), Bigfoot Biomedical (Member/Advisory Board), Dexcom (Consultant), Glooko (Member/Advisory Board), Insulet (Member/Advisory Board), Novo Nordisk (Member/Advisory Board), Zealand (Consultant).

\* *Financial, business, and organizational disclosures of the Task Force cover the year prior to publication. Disclosures prior to this time period are archived.*

\*\**Evidence-based reviews for this guideline were prepared under contract with the Endocrine Society.*

## References

- Atkins D, Best D, Briss PA, Eccles M, Falck-Ytter Y, Flottorp S, Guyatt GH, Harbour RT, Haugh MC, Henry D, Hill S, Jaeschke R, Leng G, Liberati A, Magrini N, Mason J, Middleton P, Mrukowicz J, O'Connell D, Oxman AD, Phillips B, Schunemann HJ, Edejer T, Varonen H, Vist GE, Williams JW, Jr., Zaza S. Grading quality of evidence and strength of recommendations. *BMJ*. 2004;328:1490.
- Swiglo BA, Murad MH, Schunemann HJ, Kunz R, Vigersky RA, Guyatt GH, Montori VM. A case for clarity, consistency, and helpfulness: state-of-the-art clinical practice guidelines in endocrinology using the grading of recommendations, assessment, development, and evaluation system. *J Clin Endocrinol Metab*. 2008;93:666–673.
- Guyatt GH, Schunemann HJ, Djulbegovic B, Akl EA. Guideline panels should not GRADE good practice statements. *J Clin Epidemiol*. 2015;68:597–600.
- Group TDCaCTR. The effect of intensive treatment of diabetes on the development and progression of long-term complications in insulin-dependent diabetes mellitus. The Diabetes Control and Complications Trial Research Group. *N Engl J Med* 1993;329:977–986.
- Group AC, Patel A, MacMahon S, Chalmers J, Neal B, Billot L, Woodward M, Marre M, Cooper M, Glasziou P, Grobbee D, Hamet P, Harrap S, Heller S, Liu L, Mancia G, Mogensen CE, Pan C, Poulter N, Rodgers A, Williams B, Bompont S, de Galan BE, Joshi R, Travert F. Intensive blood glucose control and vascular outcomes in patients with type 2 diabetes. *N Engl J Med*. 2008;358:2560–2572.
- Holman RR, Paul SK, Bethel MA, Matthews DR, Neil HA. 10-year follow-up of intensive glucose control in type 2 diabetes. *N Engl J Med*. 2008;359:1577–1589.
- Nathan DM, Group DER. The diabetes control and complications trial/epidemiology of diabetes interventions and complications study at 30 years: overview. *Diabetes Care*. 2014;37:9–16.
- Lachin JM, Orchard TJ, Nathan DM, Group DER. Update on cardiovascular outcomes at 30 years of the diabetes control and complications trial/epidemiology of diabetes interventions and complications study. *Diabetes Care*. 2014;37:39–43.
- Writing Group for the DERG, Orchard TJ, Nathan DM, Zinman B, Cleary P, Brillon D, Backlund JY, Lachin JM. Association between 7 years of intensive treatment of type 1 diabetes and long-term mortality. *JAMA*. 2015;313:45–53.
- American Diabetes A. 7. *Approaches to Glycemic Treatment*. *Diabetes Care*. 2015;38:S41–S48.
- FDA U. Human Factors Points to Consider for IDE Devices. Protecting and Promoting your Health [Internet Web Page]. 2015; <http://www.fda.gov/RegulatoryInformation/Guidances/ucm094531.htm>. Accessed Feb. 27, 2016.
- Implementation of treatment protocols in the Diabetes Control and Complications Trial. *Diabetes Care* 1995;18:361–376.
- Beck RW, Tamborlane WV, Bergenstal RM, Miller KM, DuBose SN, Hall CA, Network TDEC. The T1D Exchange clinic registry. *J Clin Endocrinol Metab*. 2012;97:4383–4389.
- Neithercott T. Insulin Pumps 2014. *Diabetes Forecast* 2014;
- Grunberger G, Abelseth JM, Bailey TS, Bode BW, Handelsman Y, Hellman R, Jovanovic L, Lane WS, Raskin P, Tamborlane WV, Rothermel C. Consensus statement by the American Association of Clinical Endocrinologists/American College of Endocrinology insulin pump management task force. *Endocr Pract*. 2014;20:463–489.
- Grunberger G, Bailey TS, Cohen AJ, Flood TM, Handelsman Y, Hellman R, Jovanovic L, Moghissi ES, Orzeck EA. Statement by the American Association of Clinical Endocrinologists Consensus Panel on insulin pump management. *Endocr Pract*. 2010;16:746–762.
- Excellence NifHaC. 2008; <http://www.nice.org.uk/guidance/ta151resources/guidance-continuous-insulin-infusion-for-the-treatment-of-diabetes-mellitus-pdf>.
- Pickup J, Keen H. Continuous subcutaneous insulin infusion at 25 years: evidence base for the expanding use of insulin pump therapy in type 1 diabetes. *Diabetes Care*. 2002;25:593–598.
- Misso ML, Egberts KJ, Page M, O'Connor D, Shaw J. Continuous subcutaneous insulin infusion (CSII) versus multiple insulin injections for type 1 diabetes mellitus. *Cochrane Database Syst Rev* 2010: CD005103.
- Yeh HC, Brown TT, Maruthur N, Ranasinghe P, Berger Z, Suh YD, Wilson LM, Haberl EB, Brick J, Bass EB, Golden SH. Comparative effectiveness and safety of methods of insulin delivery and glucose monitoring for diabetes mellitus: a systematic review and meta-analysis. *Ann Intern Med*. 2012;157:336–347.
- Monami M, Lamanna C, Marchionni N, Mannucci E. Continuous subcutaneous insulin infusion versus multiple daily insulin injections in type 1 diabetes: a meta-analysis. *Acta Diabetol*. 2010;47 Suppl 1:77–81.
- Jeitler K, Horvath K, Berghold A, Gratzer TW, Neeser K, Pieber TR, Siebenhofer A. Continuous subcutaneous insulin infusion versus multiple daily insulin injections in patients with diabetes mellitus: systematic review and meta-analysis. *Diabetologia*. 2008;51:941–951.
- Weissberg-Benchell J, Antisdel-Lomaglio J, Seshadri R. Insulin pump therapy: a meta-analysis. *Diabetes Care*. 2003;26:1079–1087.
- DeVries JH, Snoek FJ, Kostense PJ, Masurel N, Heine RJ, Dutch Insulin Pump Study G. A randomized trial of continuous subcutaneous insulin infusion and intensive injection therapy in type 1 diabetes for patients with long-standing poor glycemic control. *Diabetes Care*. 2002;25:2074–2080.
- Tsui E, Barnie A, Ross S, Parkes R, Zinman B. Intensive insulin therapy with insulin lispro: a randomized trial of continuous subcutaneous insulin infusion versus multiple daily insulin injection. *Diabetes Care*. 2001;24:1722–1727.
- Bolli GB, Kerr D, Thomas R, Torlone E, Sola-Gazagnes A, Vita-colonna E, Selam JL, Home PD. Comparison of a multiple daily insulin injection regimen (basal once-daily glargine plus mealtime

- lispro) and continuous subcutaneous insulin infusion (lispro) in type 1 diabetes: a randomized open parallel multicenter study. *Diabetes Care*. 2009;32:1170–1176.
27. Hirsch IB, Bode BW, Garg S, Lane WS, Sussman A, Hu P, Santiago OM, Kolaczynski JW, Insulin Aspart CMDICSG. Continuous subcutaneous insulin infusion (CSII) of insulin aspart versus multiple daily injection of insulin aspart/insulin glargine in type 1 diabetic patients previously treated with CSII. *Diabetes Care*. 2005;28:533–538.
  28. Orr CJ, Hopman W, Yen JL, Houlden RL. Long-Term Efficacy of Insulin Pump Therapy on Glycemic Control in Adults with Type 1 Diabetes Mellitus. *Diabetes Technol Ther* 2014;
  29. Pickup JC, Sutton AJ. Severe hypoglycaemia and glycaemic control in Type 1 diabetes: meta-analysis of multiple daily insulin injections compared with continuous subcutaneous insulin infusion. *Diabet Med*. 2008;25:765–774.
  30. Bruttomesso D, Crazzolara D, Maran A, Costa S, Dal Pos M, Girelli A, Lepore G, Aragona M, Iori E, Valentini U, Del Prato S, Tiengo A, Bühr A, Trevisan R, Baritussio A. In Type 1 diabetic patients with good glycaemic control, blood glucose variability is lower during continuous subcutaneous insulin infusion than during multiple daily injections with insulin glargine. *Diabet Med*. 2008;25:326–332.
  31. Pickup JC, Kidd J, Burmiston S, Yemane N. Determinants of glycaemic control in type 1 diabetes during intensified therapy with multiple daily insulin injections or continuous subcutaneous insulin infusion: importance of blood glucose variability. *Diabetes Metab Res Rev*. 2006;22:232–237.
  32. Hoogma RP, Hammond PJ, Gomis R, Kerr D, Bruttomesso D, Bouter KP, Wiefels KJ, de la Calle H, Schweitzer DH, Pfohl M, Torlone E, Krinelke LG, Bolli GB, Nations Study G. Comparison of the effects of continuous subcutaneous insulin infusion (CSII) and NPH-based multiple daily insulin injections (MDI) on glycaemic control and quality of life: results of the 5-nations trial. *Diabet Med*. 2006;23:141–147.
  33. Hanaire-Broutin H, Melki V, Bessieres-Lacombe S, Tauber JP. Comparison of continuous subcutaneous insulin infusion and multiple daily injection regimens using insulin lispro in type 1 diabetic patients on intensified treatment: a randomized study. *The Study Group for the Development of Pump Therapy in Diabetes*. *Diabetes Care*. 2000;23:1232–1235.
  34. Papargyri P, Ojeda Rodriguez S, Corrales Hernandez JJ, Mories Alvarez MT, Recio Cordova JM, Delgado Gomez M, Sanchez Marcos AI, Iglesias Lopez RA, Herrero Ruiz A, Beaulieu Oriol M, Miralles Garcia JM. An observational 7-year study of continuous subcutaneous insulin infusion for the treatment of type 1 diabetes mellitus. *Endocrinol Nutr*. 2014;61:141–146.
  35. Siegelaar SE, Holleman F, Hoekstra JB, DeVries JH. Glucose variability; does it matter? *Endocr Rev*. 2010;31:171–182.
  36. Golden SH, Brown T, Yeh HC, Maruthur N, Ranasinghe P, Berger Z, Suh Y, Wilson LM, Haberl EB, Bass EB. AHRQ Comparative Effectiveness Reviews. Methods for Insulin Delivery and Glucose Monitoring: Comparative Effectiveness. Rockville (MD): Agency for Healthcare Research and Quality (US); 2012.
  37. Colberg S. Exercising with an insulin pump. *Diabetes Self Manag* 2002;19:63–64, 67–68, 70.
  38. Yardley JE, Iscoe KE, Sigal RJ, Kenny GP, Perkins BA, Riddell MC. Insulin pump therapy is associated with less post-exercise hyperglycemia than multiple daily injections: an observational study of physically active type 1 diabetes patients. *Diabetes Technol Ther*. 2013;15:84–88.
  39. Admon G, Weinstein Y, Falk B, Weintrob N, Benzaquen H, Ofan R, Fayman G, Zigel L, Constantini N, Phillip M. Exercise with and without an insulin pump among children and adolescents with type 1 diabetes mellitus. *Pediatrics*. 2005;116:e348–355.
  40. Diabetes Research in Children Network Study G, Tsalikian E, Kollman C, Tamborlane WB, Beck RW, Fiallo-Scharer R, Fox L, Janz KF, Ruedy KJ, Wilson D, Xing D, Weinzimer SA. Prevention of hypoglycemia during exercise in children with type 1 diabetes by suspending basal insulin. *Diabetes Care*. 2006;29:2200–2204.
  41. Sharma D, Morrison G, Joseph F, Purewal TS, Weston PJ. The role of continuous subcutaneous insulin infusion therapy in patients with diabetic gastroparesis. *Diabetologia*. 2011;54:2768–2770.
  42. Barnard KD, Lloyd CE, Skinner TC. Systematic literature review: quality of life associated with insulin pump use in Type 1 diabetes. *Diabet Med*. 2007;24:607–617.
  43. Linkeschova R, Raoul M, Bott U, Berger M, Spraul M. Less severe hypoglycaemia, better metabolic control, and improved quality of life in Type 1 diabetes mellitus with continuous subcutaneous insulin infusion (CSII) therapy; an observational study of 100 consecutive patients followed for a mean of 2 years. *Diabet Med*. 2002;19:746–751.
  44. Thomas RM, Aldibbiat A, Griffin W, Cox MA, Leech NJ, Shaw JA. A randomized pilot study in Type 1 diabetes complicated by severe hypoglycaemia, comparing rigorous hypoglycaemia avoidance with insulin analogue therapy, CSII or education alone. *Diabet Med*. 2007;24:778–783.
  45. Berthe E, Lireux B, Coffin C, Goulet-Salmon B, Houlbert D, Boutreux S, Fradin S, Reznik Y. Effectiveness of intensive insulin therapy by multiple daily injections and continuous subcutaneous insulin infusion: a comparison study in type 2 diabetes with conventional insulin regimen failure. *Horm Metab Res*. 2007;39:224–229.
  46. Edelman SV, Bode BW, Bailey TS, Kipnes MS, Brunelle R, Chen X, Frias JP. Insulin pump therapy in patients with type 2 diabetes safely improved glycaemic control using a simple insulin dosing regimen. *Diabetes Technol Ther*. 2010;12:627–633.
  47. Peyrot M, Rubin RR, Chen X, Frias JP. Associations between improved glucose control and patient-reported outcomes after initiation of insulin pump therapy in patients with type 2 diabetes mellitus. *Diabetes Technol Ther*. 2011;13:471–476.
  48. Pickup JC, Renard E. Long-acting insulin analogs versus insulin pump therapy for the treatment of type 1 and type 2 diabetes. *Diabetes Care*. 2008;31 Suppl 2:S140–145.
  49. Raskin P, Bode BW, Marks JB, Hirsch IB, Weinstein RL, McGill JB, Peterson GE, Mudaliar SR, Reinhardt RR. Continuous subcutaneous insulin infusion and multiple daily injection therapy are equally effective in type 2 diabetes: a randomized, parallel-group, 24-week study. *Diabetes Care*. 2003;26:2598–2603.
  50. Herman WH, Ilag LL, Johnson SL, Martin CL, Sinding J, Al Harthi A, Plunkett CD, LaPorte FB, Burke R, Brown MB, Halter JB, Raskin P. A clinical trial of continuous subcutaneous insulin infusion versus multiple daily injections in older adults with type 2 diabetes. *Diabetes Care*. 2005;28:1568–1573.
  51. Wainstein J, Metzger M, Boaz M, Minuchin O, Cohen Y, Yaffe A, Yerushalmy Y, Raz I, Harman-Boehm I. Insulin pump therapy vs. multiple daily injections in obese Type 2 diabetic patients. *Diabet Med*. 2005;22:1037–1046.
  52. Reznik Y, Cohen O, Aronson R, Conget I, Runzis S, Castaneda J, Lee SW, Op TmSG. Insulin pump treatment compared with multiple daily injections for treatment of type 2 diabetes (OpT2mise): a randomised open-label controlled trial. *Lancet* 2014;384:1265–1272.
  53. Inzucchi SE, Bergenstal RM, Buse JB, Diamant M, Ferrannini E, Nauck M, Peters AL, Tsapas A, Wender R, Matthews DR. Management of hyperglycemia in type 2 diabetes, 2015: a patient-centered approach: update to a position statement of the American Diabetes Association and the European Association for the Study of Diabetes. *Diabetes Care*. 2015;38:140–149.
  54. Lane WS. Use of U-500 regular insulin by continuous subcutaneous insulin infusion in patients with type 2 diabetes and severe insulin resistance. *Endocr Pract*. 2006;12:251–256.
  55. Lane WS, Weinrib SL, Rappaport JM, Hale CB, Farmer LK, Lane RS. The effect of long-term use of U-500 insulin via continuous subcutaneous infusion on durability of glycaemic control and weight in obese, insulin-resistant patients with type 2 diabetes. *Endocr Pract*. 2013;19:196–201.
  56. Reutrakul S, Wroblewski K, Brown RL. Clinical use of U-500 reg-

- ular insulin: review and meta-analysis. *J Diabetes Sci Technol*. 2012; 6:412–420.
57. Bulchandani DG, Konrady T, Hamburg MS. Clinical efficacy and patient satisfaction with U-500 insulin pump therapy in patients with type 2 diabetes. *Endocr Pract*. 2007;13:721–725.
  58. Crasto W, Jarvis J, Hackett E, Nayyar V, McNally PG, Davies MJ, Lawrence IG. Insulin U-500 in severe insulin resistance in type 2 diabetes mellitus. *Postgrad Med J*. 2009;85:219–222.
  59. Jones P, Idris I. The use of U-500 regular insulin in the management of patients with obesity and insulin resistance. *Diabetes Obes Metab*. 2013;15:882–887.
  60. Boyle ME, Seifert KM, Beer KA, Apsey HA, Nassar AA, Littman SD, Magallanez JM, Schlinkert RT, Stearns JD, Hovan MJ, Cook CB. Guidelines for application of continuous subcutaneous insulin infusion (insulin pump) therapy in the perioperative period. *J Diabetes Sci Technol*. 2012;6:184–190.
  61. Morviducci L, Di Flaviani A, Lauria A, Pitocco D, Pozzilli P, Suraci C, Frontoni S. Continuous subcutaneous insulin infusion (CSII) in inpatient setting: unmet needs and the proposal of a CSII unit. *Diabetes Technol Ther*. 2011;13:1071–1074.
  62. Houlden RL, Moore S. In-hospital management of adults using insulin pump therapy. *Can J Diabetes*. 2014;38:126–133.
  63. Lansang MC, Modic MB, Sauvey R, Lock P, Ross D, Combs P, Kennedy L. Approach to the adult hospitalized patient on an insulin pump. *J Hosp Med*. 2013;8:721–727.
  64. Cook CB, Beer KA, Seifert KM, Boyle ME, Mackey PA, Castro JC. Transitioning insulin pump therapy from the outpatient to the inpatient setting: a review of 6 years' experience with 253 cases. *J Diabetes Sci Technol*. 2012;6:995–1002.
  65. Boyle ME, Seifert KM, Beer KA, Mackey P, Schlinkert RT, Stearns JD, Cook CB. Insulin pump therapy in the perioperative period: a review of care after implementation of institutional guidelines. *J Diabetes Sci Technol*. 2012;6:1016–1021.
  66. Chamberlain JJ, Rhinehart AS, Shaefer CF, Jr., Neuman A. Diagnosis and Management of Diabetes: Synopsis of the 2016 American Diabetes Association Standards of Medical Care in Diabetes. *Ann Intern Med* 2016;
  67. Scheiner G, Sobel RJ, Smith DE, Pick AJ, Kruger D, King J, Green K. Insulin pump therapy: guidelines for successful outcomes. *Diabetes Educ* 2009; 35 Suppl 2:29S–41S; quiz 28S, 42S–43S
  68. Clements M, Matuleviciene V, Attvall S, Ekelund M, Pivodic A, Dahlqvist S, Fahlen M, Haraldsson B, Lind M. Predicting the Effectiveness of Insulin Pump Therapy on Glycemic Control in Clinical Practice: A Retrospective Study of Patients with Type 1 Diabetes from 10 Outpatient Diabetes Clinics in Sweden over 5 Years. *Diabetes Technol Ther* 2014;
  69. Ziegler R, Tubili C, Chico A, Guerci B, Lundberg E, Borchert M, Loffler A, Bloethner S, Weissmann J, Pflutzner A. ProAct study: new features of insulin pumps improve diabetes management and glycemic control in patients after transition of continuous subcutaneous insulin infusion systems. *Diabetes Technol Ther*. 2013;15:738–743.
  70. Pinhas-Hamiel O, Tzadok M, Hirsh G, Boyko V, Graph-Barel C, Lerner-Geva L, Reichman B. The impact of baseline hemoglobin A1c levels prior to initiation of pump therapy on long-term metabolic control. *Diabetes Technol Ther*. 2010;12:567–573.
  71. Nixon R, Folwell R, Pickup JC. Variations in the quality and sustainability of long-term glycaemic control with continuous subcutaneous insulin infusion. *Diabet Med*. 2014;31:1174–1177.
  72. Grant P, Dworakowska D, DeZoysa N, Barnes D. The impact of anxiety and depression on patients within a large type 1 diabetes insulin pump population. An observational study. *Diabetes Metab*. 2013;39:439–444.
  73. Matejko B, Skupien J, Mrozinska S, Grzanka M, Cyganek K, Kiec-Wilk B, Malecki MT, Klupa T. Factors associated with glycemic control in adult type 1 diabetes patients treated with insulin pump therapy. *Endocrine* 2014;
  74. Matejko B, Cyganek K, Katra B, Galicka-Latala D, Grzanka M, Malecki MT, Klupa T. Insulin pump therapy is equally effective and safe in elderly and young type 1 diabetes patients. *Rev Diabet Stud*. 2011;8:254–258.
  75. Klupa T, Benbenek-Klupa T, Malecki M, Szalecki M, Sieradzki J. Clinical usefulness of a bolus calculator in maintaining normoglycaemia in active professional patients with type 1 diabetes treated with continuous subcutaneous insulin infusion. *J Int Med Res*. 2008; 36:1112–1116.
  76. Rankin D, Cooke DD, Clark M, Heller S, Elliott J, Lawton J. How and why do patients with Type 1 diabetes sustain their use of flexible intensive insulin therapy? A qualitative longitudinal investigation of patients' self-management practices following attendance at a Dose Adjustment for Normal Eating (DAFNE) course. *Diabet Med*. 2011;28:532–538.
  77. Zisser H, Robinson L, Bevier W, Dassau E, Ellingsen C, Doyle FJ, Jovanovic L. Bolus calculator: a review of four "smart" insulin pumps. *Diabetes Technol Ther*. 2008;10:441–444.
  78. Gross TM, Kayne D, King A, Rother C, Juth S. A bolus calculator is an effective means of controlling postprandial glycemia in patients on insulin pump therapy. *Diabetes Technol Ther*. 2003;5:365–369.
  79. Lepore G, Dodesini AR, Nosari I, Scaranna C, Corsi A, Trevisan R. Bolus calculator improves long-term metabolic control and reduces glucose variability in pump-treated patients with Type 1 diabetes. *Nutr Metab Cardiovasc Dis*. 2012;22:e15–16.
  80. Quiros C, Patrascioiu I, Gimenez M, Vinagre I, Vidal M, Jansa M, Conget I. Assessment of use of specific features of subcutaneous insulin infusion systems and their relationship to metabolic control in patients with type 1 diabetes. *Endocrinol Nutr*. 2014;61:318–322.
  81. Ziegler R, Cavan DA, Cranston I, Barnard K, Ryder J, Vogel C, Parkin CG, Koehler W, Vesper I, Petersen B, Schweitzer MA, Wagner RS. Use of an insulin bolus advisor improves glycemic control in multiple daily insulin injection (MDI) therapy patients with suboptimal glycemic control: first results from the ABACUS trial. *Diabetes Care*. 2013;36:3613–3619.
  82. Cavan DA, Ziegler R, Cranston I, Barnard K, Ryder J, Vogel C, Parkin CG, Koehler W, Vesper I, Petersen B, Wagner RS. Automated bolus advisor control and usability study (ABACUS): does use of an insulin bolus advisor improve glycaemic control in patients failing multiple daily insulin injection (MDI) therapy? [NCT01460446]. *BMC Fam Pract*. 2012;13:102.
  83. Klonoff DC. The current status of bolus calculator decision-support software. *J Diabetes Sci Technol*. 2012;6:990–994.
  84. Jayasekara RS, Munn Z, Lockwood C. Effect of educational components and strategies associated with insulin pump therapy: a systematic review. *Int J Evid Based Healthc*. 2011;9:346–361.
  85. Bell KJ, Smart CE, Steil GM, Brand-Miller JC, King B, Wolpert HA. Impact of fat, protein, and glycemic index on postprandial glucose control in type 1 diabetes: implications for intensive diabetes management in the continuous glucose monitoring era. *Diabetes Care*. 2015;38:1008–1015.
  86. Smart CE, Evans M, O'Connell SM, McElduff P, Lopez PE, Jones TW, Davis EA, King BR. Both dietary protein and fat increase postprandial glucose excursions in children with type 1 diabetes, and the effect is additive. *Diabetes Care*. 2013;36:3897–3902.
  87. Wolpert HA, Atakov-Castillo A, Smith SA, Steil GM. Dietary fat acutely increases glucose concentrations and insulin requirements in patients with type 1 diabetes: implications for carbohydrate-based bolus dose calculation and intensive diabetes management. *Diabetes Care*. 2013;36:810–816.
  88. Chase HP, Saib SZ, MacKenzie T, Hansen MM, Garg SK. Postprandial glucose excursions following four methods of bolus insulin administration in subjects with type 1 diabetes. *Diabet Med*. 2002; 19:317–321.
  89. Jones SM, Quarry JL, Caldwell-McMillan M, Mauger DT, Gabbay RA. Optimal insulin pump dosing and postprandial glycemia following a pizza meal using the continuous glucose monitoring system. *Diabetes Technol Ther*. 2005;7:233–240.



90. Lee SW, Cao M, Sajid S, Hayes M, Choi L, Rother C, de Leon R. The dual-wave bolus feature in continuous subcutaneous insulin infusion pumps controls prolonged post-prandial hyperglycaemia better than standard bolus in Type 1 diabetes. *Diabetes Nutr Metab*. 2004; 17:211–216.
91. Tamborlane WV, Beck RW, Bode BW, Buckingham B, Chase HP, Clemons R, Fiallo-Scharer R, Fox LA, Gilliam LK, Hirsch IB, Huang ES, Kollman C, Kowalski AJ, Laffel L, Lawrence JM, Lee J, Mauras N, O'Grady M, Ruedy KJ, Tansey M, Tsalikian E, Weinzimer S, Wilson DM, Wolpert H, Wysocki T, Xing D. Continuous glucose monitoring and intensive treatment of type 1 diabetes. *N Engl J Med*. 2008;359:1464–1476.
92. Deiss D, Bolinder J, Riveline JP, Battelino T, Bosi E, Tubiana-Rufi N, Kerr D, Phillip M. Improved glycemic control in poorly controlled patients with type 1 diabetes using real-time continuous glucose monitoring. *Diabetes Care*. 2006;29:2730–2732.
93. O'Connell MA, Donath S, O'Neal DN, Colman PG, Ambler GR, Jones TW, Davis EA, Cameron FJ. Glycaemic impact of patient-led use of sensor-guided pump therapy in type 1 diabetes: a randomised controlled trial. *Diabetologia*. 2009;52:1250–1257.
94. Buckingham B, Beck RW, Tamborlane WV, Xing D, Kollman C, Fiallo-Scharer R, Mauras N, Ruedy KJ, Tansey M, Weinzimer SA, Wysocki T. Continuous glucose monitoring in children with type 1 diabetes. *J Pediatr* 2007;151:388–393, 393.e381–382.
95. Beck RW, Hirsch IB, Laffel L, Tamborlane WV, Bode BW, Buckingham B, Chase P, Clemons R, Fiallo-Scharer R, Fox LA, Gilliam LK, Huang ES, Kollman C, Kowalski AJ, Lawrence JM, Lee J, Mauras N, O'Grady M, Ruedy KJ, Tansey M, Tsalikian E, Weinzimer SA, Wilson DM, Wolpert H, Wysocki T, Xing D. The effect of continuous glucose monitoring in well-controlled type 1 diabetes. *Diabetes Care*. 2009;32:1378–1383.
96. Price D. Multiple Daily Injections and Continuous Glucose Monitoring in Diabetes <https://clinicaltrials.gov/ct2/show/NCT02282397?term=NCT02282397&rank=1>.
97. Bode B, Beck RW, Xing D, Gilliam L, Hirsch I, Kollman C, Laffel L, Ruedy KJ, Tamborlane WV, Weinzimer S, Wolpert H. Sustained benefit of continuous glucose monitoring on A1C, glucose profiles, and hypoglycemia in adults with type 1 diabetes. *Diabetes Care*. 2009;32:2047–2049.
98. Hirsch IB, Bode BW, Childs BP, Close KL, Fisher WA, Gavin JR, Ginsberg BH, Raine CH, Verderese CA. Self-Monitoring of Blood Glucose (SMBG) in insulin- and non-insulin-using adults with diabetes: consensus recommendations for improving SMBG accuracy, utilization, and research. *Diabetes Technol Ther*. 2008;10:419–439.
99. Sequeira PA, Montoya L, Ruelas V, Xing D, Chen V, Beck R, Peters AL. Continuous glucose monitoring pilot in low-income type 1 diabetes patients. *Diabetes Technol Ther*. 2013;15:855–858.
100. Battelino T, Phillip M, Bratina N, Nimri R, Oskarsson P, Bolinder J. Effect of continuous glucose monitoring on hypoglycemia in type 1 diabetes. *Diabetes Care*. 2011;34:795–800.
101. Vigersky RA, Fonda SJ, Chellappa M, Walker MS, Ehrhardt NM. Short- and long-term effects of real-time continuous glucose monitoring in patients with type 2 diabetes. *Diabetes Care*. 2012;35:32–38.
102. Ehrhardt NM, Chellappa M, Walker MS, Fonda SJ, Vigersky RA. The effect of real-time continuous glucose monitoring on glycemic control in patients with type 2 diabetes mellitus. *J Diabetes Sci Technol*. 2011;5:668–675.
103. Yoo HJ, An HG, Park SY, Ryu OH, Kim HY, Seo JA, Hong EG, Shin DH, Kim YH, Kim SG, Choi KM, Park IB, Yu JM, Baik SH. Use of a real time continuous glucose monitoring system as a motivational device for poorly controlled type 2 diabetes. *Diabetes Res Clin Pract*. 2008;82:73–79.
104. Gandhi GY, Kovalaske M, Kudva Y, Walsh K, Elamin MB, Beers M, Coyle C, Goalen M, Murad MS, Erwin PJ, Corpus J, Montori VM, Murad MH. Efficacy of continuous glucose monitoring in improving glycemic control and reducing hypoglycemia: a systematic review and meta-analysis of randomized trials. *J Diabetes Sci Technol*. 2011;5:952–965.
105. Cosson E, Hamo-Tchatchouang E, Dufaitre-Patouraux L, Attali JR, Paries J, Schaepelynck-Belicar P. Multicentre, randomised, controlled study of the impact of continuous sub-cutaneous glucose monitoring (GlucoDay) on glycaemia control in type 1 and type 2 diabetes patients. *Diabetes Metab*. 2009;35:312–318.
106. Allen NA, Fain JA, Braun B, Chipkin SR. Continuous glucose monitoring counseling improves physical activity behaviors of individuals with type 2 diabetes: A randomized clinical trial. *Diabetes Res Clin Pract*. 2008;80:371–379.
107. Munshi MN, Segal AR, Suhl E, Staum E, Desrochers L, Sternthal A, Giusti J, McCartney R, Lee Y, Bonsignore P, Weinger K. Frequent hypoglycemia among elderly patients with poor glycemic control. *Arch Intern Med*. 2011;171:362–364.
108. Hunger-Dathe W, Braun A, Muller UA, Schiel R, Femerling M, Risse A. Insulin pump therapy in patients with Type 1 diabetes mellitus: results of the Nationwide Quality Circle in Germany (ASD) 1999–2000. *Exp Clin Endocrinol Diabetes*. 2003;111:428–434.
109. Training in flexible, intensive insulin management to enable dietary freedom in people with type 1 diabetes: dose adjustment for normal eating (DAFNE) randomised controlled trial. *BMJ* 2002; 325:746.
110. Taylor J, Cottrell C, Chatterton H, Hill J, Hughes R, Wohlgenuth C, Holt RI. Identifying risk and preventing progression to Type 2 diabetes in vulnerable and disadvantaged adults: a pragmatic review. *Diabet Med*. 2013;30:16–25.
111. Monami M, Lamanna C, Marchionni N, Mannucci E. Continuous subcutaneous insulin infusion versus multiple daily insulin injections in type 2 diabetes: a meta-analysis. *Exp Clin Endocrinol Diabetes*. 2009;117:220–222.
112. Bergenstal RM, Tamborlane WV, Ahmann A, Buse JB, Dailey G, Davis SN, Joyce C, Peoples T, Perkins BA, Welsh JB, Willi SM, Wood MA. Effectiveness of sensor-augmented insulin-pump therapy in type 1 diabetes. *N Engl J Med*. 2010;363:311–320.
113. Powers MA, Marrero DG. Response to Comment on Powers et al. Diabetes Self-management Education and Support in Type 2 Diabetes: A Joint Position Statement of the American Diabetes Association, the American Association of Diabetes Educators, and the Academy of Nutrition and Dietetics. *Diabetes Care* 2015;38: 1372–1382. *Diabetes Care*. 2016;39:e17.
114. Evert A, Trence D, Catton S, Huynh P. Continuous glucose monitoring technology for personal use: an educational program that educates and supports the patient. *Diabetes Educ* 2009;35:565–567, 571–563, 577–580.
115. Gilliam LK, Hirsch IB. Practical aspects of real-time continuous glucose monitoring. *Diabetes Technol Ther*. 2009;11 Suppl 1:S75–82.
116. Bergenstal RM, Klonoff DC, Garg SK, Bode BW, Meredith M, Slover RH, Ahmann AJ, Welsh JB, Lee SW, Kaufman FR. Threshold-based insulin-pump interruption for reduction of hypoglycemia. *N Engl J Med*. 2013;369:224–232.
117. Russell SJ, El-Khatib FH, Sinha M, Magyar KL, McKeon K, Gergen LG, Balliro C, Hillard MA, Nathan DM, Damiano ER. Out-patient glycemic control with a bionic pancreas in type 1 diabetes. *N Engl J Med*. 2014;371:313–325.
118. Thabit H, Tauschmann M, Allen JM, Leelarathna L, Hartnell S, Wilinska ME, Acerini CL, Dellweg S, Benesch C, Heinemann L, Mader JK, Holzer M, Kojzar H, Exall J, Yong J, Pichierrri J, Barnard KD, Kollman C, Cheng P, Hindmarsh PC, Campbell FM, Arnolds S, Pieber TR, Evans ML, Dunger DB, Hovorka R. Home Use of an Artificial Beta Cell in Type 1 Diabetes. *N Engl J Med*. 2015;373: 2129–2140.
119. Heinemann L, Fleming GA, Petrie JR, Holl RW, Bergenstal RM, Peters AL. Insulin pump risks and benefits: a clinical appraisal of pump safety standards, adverse event reporting, and research needs: a joint statement of the European Association for the Study



of Diabetes and the American Diabetes Association Diabetes Technology Working Group. *Diabetes Care*. 2015;38:716–722.